

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Solan H.P 174103

CIN: U71200HP2023SGC010333

Phone: - +91 93563 01674, E-mail: - himpharmalab@gmail.com

Ref. Tender Notice No. HPTLL_2025_03

DATED 05.03.2025

Please find enclosed herewith the following:		Page No.
1.	Tender Notice No. HPTLL_2025_03 & Schedule for Opening of the Tender	2
2.	Annexure-I (Detail of Equipment, Tender Fee & EMD)	3
3.	Important Note	4
4.	Eligibility Criteria	5
5.	Instruction to Tenderers	6-8
6.	Tender Evaluation	9-10
7.	Terms & Conditions	11-12
8.	Acceptance of terms and conditions	13
9.	Performa for performance statement	14
10.	Annexure 'A' (Specifications of the Equipment)	15-25
11.	Annexure 'B' & 'C' (Format for Performance Bond/ Guarantee & Format for Performance Bond (Bank Guarantee)	26-29
12.	Annexure – 'D' (Technical Bid Form)	30
13.	Annexure- 'E' (Format for Manufacturer's Authorization Form)	31
14.	Annexure –'F' (Declaration Regarding Blacklisting/ Debarring for taking part in Tender)	32
15.	Annexure- 'G' (Certificate of Warranty)	33
16.	Annexure- 'H' (Financial Bid Form)	34
17.	Annexure- 'I' (Consignee Receipt Certificate)	35
18.	Annexure- 'J' (Proforma of Final Acceptance Certificate by the Consignee)	36-37

CEO

Himachal Pharma Testing Lab Limited

E-mail: himpharmalab@gmail.com

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P 174103

CIN: U71200HP2023SGC010333

Phone: - +91 93563 01674, E-mail: - himpfarmalab@gmail.com

TENDER NOTICE

Date: 05.03.2025

HIMACHAL PHARMA TESTING LAB LIMITED, BADDI, HP (a company incorporated under the Companies Act, 2013) invites sealed tenders for **Procurement, supply, Installation and Commission of Lab Equipment's** for the following items for its Testing Laboratory at Himachal Pharma Testing Lab Ltd, located at EPIP, Phase-1, Jharmajri, Baddi, Distt. Solan (H.P). 174103

I	Downloading & Submission of tender/bids	Start Date: 05.03.2025, Wednesday at 11:00 AM
II	Last date of submission of Physical bids, Tender Fee and EMD	End Date: 19.03.2025, Wednesday at 11:00 AM
III	Pre bid date Online	11.03.2025, Tuesday
IV	Opening of Technical Bid (offline)	20.03.2025, Thursday at 11:00 AM

Detailed Terms and Conditions are available in tender document. The bid document can be downloaded from the www.hptll.in

Complete tender document is available for reference purposes on our website www.hptll.in

CEO, Himachal Pharma Testing Lab Limited

ANNEXURE: I

Sr. No.	Tender ID	Description	Tender Fee	Total estimated cost (In Rs.)	EMD (In Rs)
1	HPTLL_2025_03	GCMS with autosampler and HS and ECD Detector: 1 Nos, Ultraviolet/ Visible Spectrophotometer: 1 Nos.	5,000/-	1,77,04,430/-	3,54,089/-

*Exemption of Tender Fee and EMD will only be given to MSE (Micro & Small Enterprises)

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail: himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Important Note

1. All corrigenda, addenda, amendments and clarifications regarding this tender document will be uploaded on the website www.hptll.in and not in the newspaper; Bidders shall keep themselves updated with all such developments.
2. In case, the last date of receipt/opening of bids falls on holiday, the bids shall be receipt/opened on the next working day at same time.
3. In case, the last date of submission of EMD & Tender fee falls on holiday, the EMD & Tender fee shall be submitted on the next working day at same time.
4. Tenderer who have downloaded the tender documents form from the Himachal Pharma Testing Lab Limited website, shall submit a declaration along with tender document that I/We have downloaded the Tender Form from the Himachal Pharma Testing Lab Limited website www.hptll.in and I/we have not tempered /modified the tender form in any manner. In case, if the same is found to be tempered/modified in any manner, I/we understand that my/our tender will be summarily rejected and I/we are liable to be banned from doing business with Himachal Pharma Testing Lab Limited.
5. **Tender fee of Rs. 5,000/- (Rs. Five Thousand Only) (Non- refundable) in the form of DD in favour of “Himachal Pharma Testing Lab Limited”**
6. **EMD at 2% (refundable) in the form of DD in favour of “Himachal Pharma Testing Lab Limited” payable at Panchkula.**
7. **Both EMD and Tender fee are be submitted as per dates mentioned in schedule, failing which e-bids will not be considered.**



All the bidders are required to submit the Tender Fee and EMD as per requirement of tender document failing which bids received straightway rejected and bid will be treated invalid.



Note: If the bidder inadvertently or otherwise submits the quoted rates in the technical bid, the bid will be straightway rejected and treated invalid.

Bidder must submit a scanned copy (duly signed and stamped) regarding terms & conditions as per our tender documents along-with make/model, specifications, bill of quantity as per required equipment in the technical bid for examine the bid as per our Himachal Pharma Testing Lab Limited tender documents. It is noted that no rate should be depicted in the letter head.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

ELIGIBILITY CRITERIA:

1. The bidder must have to submit experience certificate of last five financial years regarding satisfactory supply/work execution of similar nature to Govt/semi govt/ corporate organization or state/Centre PSU/Public/Private Companies/ preferably in the state of HP of at least 80% of Total estimated cost. (Attached: Proforma for Performance Statement)
2. The bidder has to submit an **earnest money of Rs. 3,54,089/- (Rupees Three Lakh Fifty-Four Thousand Eighty-Nine Rupees Only)** in the shape of DD only in favour of Himachal Pharma Testing Lab Limited. EMD of unsuccessful bidders will be returned after finalizing the tender. In case any bidder withdraws his offer after submission of tender the undersigned may forfeit the EMD. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. MSME exemptions are allowed only to Original Equipment Manufacturers (OEMs), not to dealers.
3. The Technical bid must conform to the specification as uploaded on our website i.e. www.hptll.in
4. Prospective Bidders must be manufacturer or authorized dealers of the quoted item(s) and must submit hard bound Bid Documents all pages duly serially numbered. Spiral or loose bidding shall not be entertained
5. The bidder should not have been debarred and/ or blacklisted by any Central Government/ or any State Government Department(s). This must be supported by an affidavit as per format given in Annexure-“F”.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Instructions to Tenderer

1. The Tender will be accepted only in physical form. The bidders shall have to submit their bids at the office of CEO, Himachal Pharma Testing Lab Limited, EPIP-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Distt Solan H.P. 174103.
2. Bids are to be submitted physical and opened offline at Himachal Pharma Testing Lab Limited, Baddi as per time given failing which no bid will be considered.
3. Before submission of bids, bidders must ensure that copies of all the necessary/relevant documents have been submitted with the bid which should be duly signed and stamped. The duly signed and stamped copies of Terms & Conditions of the tender, and other documents of the Tender & Annexures must be submitted, failing which their bids may be rejected.
4. Himachal Pharma Testing Lab Limited, will not be responsible for any delay in submission of bids due to any reason whatsoever.
5. The bidder must have to submit experience certificate of last five financial years regarding satisfactory supply/work execution of similar nature to Govt/semi govt/ corporate organization or state/Centre PSU/Public/Private Companies/ preferably in the state of HP of at least 80% of Total estimated cost.
6. The Bidder Should have an average annual financial turnover not less than Rs. **50 Crores** during the last three years ending 31st March 2024. Group turnover of associate/sister entities will also be considered for the purpose of average annual financial turnover. (Audited Balance sheet to be attached).
7. Bidders should also submit the copies of Tender fees/EMD as specified in the tender documents along with technical documents. **EMD** in the form of a Demand Draft in favour of the Himachal Pharma Testing Lab Limited, payable at Panchkula (refundable separate) and **Tender Fee** in the form of a Demand Draft in favour of Himachal Pharma Testing Lab Limited, payable at Panchkula (Non- refundable separate) should also be submitted in physical form to the following address as per scheduled time given for physical submission of EMD and Tender fee. The Envelope should be super-scribed as **“EMD and Tender Fee for Tender for Supply of Lab Equipment's” at Himachal Pharma Testing Lab Limited**. The envelopes inside main envelope should contain the following:

Envelope	Marked on the Cover	Contents of Envelope
Main Envelop	Subject of Tender	Envelop-I+ Envelop-II

Envelop-I	Technical Bid	Should contain Tender Document Cost in original. Should contain EMD in original. Should contain Tender Form-Technical Bid with all pages of Term and Conditions with signature on each page.
Envelop-II	Financial Bid	Rates should be quoted in the prescribed Tender Form-Financial Bid Format only with signature on each page.

Address for submission of bid:

**Kind Attention- Chief Executive Officer (CEO)
Himachal Pharma Testing Lab Limited
EPIP, Phase-1, Near BBNDA Office, Jharmajri
Tehsil Baddi, Distt Solan H.P 174103**

8. The details of EMD specified in the tender document should be same as submitted. Otherwise tender will be rejected summarily.
9. The conditional bids shall not be considered and will be out rightly rejected.
10. The Financial Bid through tendering process shall be opened only of those bidders, who will qualify in the technical bid and approved by the Purchase Committee/Technical Experts. The date, time & place of opening of the financial bid(s) will be intimated in due course of time.
11. At any time prior to the deadline for submission of bid, Himachal Pharma Testing Lab Limited may, for any reason, whether at its own initiative or in response to clarification(s) requested by a prospective tenderer(s), modify the tender document(s) by issuance of an amendment.
12. The amendment will be uploaded on Himachal Pharma Testing Lab Limited website i.e. www.hptll.in only. In order to provide reasonable time to prospective tenderer(s), for preparing their bid as per amendment, Himachal Pharma Testing Lab Limited may, at its discretion extend the deadline for the submission of tender.
13. **The supplier must submit the original manuals / catalogue and Make/Model of the Equipment /Item otherwise bid is liable to be rejected.**
14. Himachal Pharma Testing Lab Limited is not liable to pay any interest on EMD. Earnest money deposit shall be forfeited, if the tenderer, withdraws its bid during the period of tender validity. The Earnest Money Deposit of the tenderer, whose tender has been accepted, will be returned on the submission of **performance security @ 5% of the total value of the offer. The performance security will be kept till the warranty period of one year. The warranty period will start from the date of satisfactory installation of the Equipment /Item duly given by the concerned department.** Earnest Money Deposit of the successful tenderer shall be forfeited, if it refuses or neglects to execute the contract or fails to furnish the required performance security within the time frame as specified by Himachal Pharma Testing Lab Limited. The EMD(s) of other Bidder(s) whose offer are found according to required specifications will be released after finalization of Technical and Financial Lowest Bid(s). The Bidder should submit the Financial Bid as per the format given in Annexure **"H"**
15. The Format of Performance Bank Guarantee bond or Performance Bank Guarantee issued by the bank as per the format given in **Annexure "B" & "C"**.

16. Delivery time is the essence of the contract i.e. 60 days.
17. Nearest specifications/better specifications can be considered. In case of deviation, complete justification should be furnished with proper documents.
18. The Procurement-cum Evaluation Committee of Himachal Pharma Testing Lab Limited may accept a tender in part or whole of the quantity offered, reject any tender without assigning any reasons and may not accept the lowest bidder. Further in case of any doubt/dispute, the decision of the Procurement-cum Evaluation Committee of Director of the Himachal Pharma Testing Lab Limited shall be final.
19. The offer shall be kept valid for minimum 90 days.
20.
 - a) Either the agent/ dealer on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender.
 - b) If an agent/ dealer submits bid on behalf of the Principal/OEM, the same agent /Dealer shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product.
 - c) All offers other than those from the Principal/OEM should be supported by an authority letter from the manufacturer authorizing the dealer /supplier to tender on their behalf as per **Annexure-E**. In case of manufacturer, a certificate or a copy thereof to the effect that the bidder is a manufacturer of the Equipment/Item must be accompanied with the technical bid.
21. The supplier will be responsible till the entire stores contracted for, arrive in good condition at destination.
22. If any information furnished by the bidder is, at any stage found to be incorrect/false/fabricated, Himachal Pharma Testing Lab Limited shall have the absolute right to forfeit the EMD, warranty/performance guarantees or/and security deposits, in addition to cancellation of contract, and in accordance with law, such other actions may be taken like black-listing of the bidder etc.

**Read and accepted.
Tenderer)**

(Signature & Stamp of

TENDER EVALUATION

Himachal Pharma Testing Lab Limited will evaluate all the proposals to determine whether these are complete in all respects as specified in the tender document. Evaluation of the proposals shall be done in two stages as:

(a) Stage – I (Technical Evaluation):

- i) Himachal Pharma Testing Lab Limited shall evaluate the technical bid(s) to determine the following like the bid qualifies the essential eligibility criteria or not, the tenderer has submitted the EMD & Tender fee or not, any computation errors have been made or not, all the documents have been properly filled or otherwise, all the documents have been submitted with technical bid or not, the specifications, Make/Model, Catalogue of quoted Equipment /Item are as per requirement tender specifications or not, Authorization of Dealer / Distributor/ Exclusive Agent certificate from manufacturer is in order or not, Sales & service policy of equipment / item during warranty period and after warranty period will also be seen, location of their authorized service centre will also be seen for evaluation etc.

After evaluation of technical bid(s), a list of the qualifying tenderer (s)/ bidder

s) shall be made. Short-listed tenderer(s) will be informed of the date, time and place of opening of financial bid(s) and they may attend or depute their authorized representative/s to attend the schedule of opening of financial bid(s) on the scheduled date and time, if they wish to do so. The representative(s) should have a letter of authority to attend the price bid(s) opening event.

- ii) Bidders should offer the rates as per **Annexure H**. Detailed bill of material/quantity is also to be provided along with the price breakup of each item as per requirement of the tender specification of the equipment.

Note: The quoted amount as filled in the Annexure(s) of financial bid and detailed bill of material/quantity provided with price break up of each item in the financial bid should be tallied and both must be same, otherwise bid will be treated invalid.

Read and accepted.

(Signature & Stamp of Tenderer)

ARBITRATION:

In case of any dispute or difference arising out in connection with the tender conditions/job order/Contract, Himachal Pharma Testing Lab Limited and the Seller/Service Provider will address the dispute/difference for a mutual resolution failing which, the matter shall be referred for arbitration to a sole Arbitrator to be appointed by the HPTLL.

The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Nalagarh, Distt Solan H.P. only. The decision of the Arbitrator shall be final and binding on both the parties.

JURISDICTION:

The courts at Nalagarh alone will have the jurisdiction to trial any matter, dispute or reference between parties arising out of this tender / contract. It is specifically agreed that no court outside and other than Nalagarh Court shall have jurisdiction in the matter.

Read and accepted.

(Signature & Stamp of Tenderer)

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

(TERMS AND CONDITIONS (FOR THE SUPPLY OF GOODS, EQUIPMENT /ITEM)

1. Rate should be quoted F.O.R Himachal Pharma Testing Lab Limited, Baddi, H.P. and In INR only.
2. GST or any other chargeable duty where applicable must be specifically mentioned in the financial quote, failing which no tax or duty will be allowed at subsequent stage.
3. All items shall be indicated both in words as well as in figures. If there is difference between amount quoted in words and figures, amount quoted in words shall prevail.
4. **Payment:** Payment shall be made subject to recoveries, if any, by way of liquidated damages or any other charges as per terms & conditions of contract in the following manner.

A) Payment for Domestic Goods or Foreign origin located within India.

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) On delivery:

80 % payment of the contract price shall be made on receipt of goods in good condition and upon the submission of the following documents:

- (i) Four copies of supplier's invoice indicating Bill to Himachal Pharma Testing Lab Limited and Ship to Place of supply, showing contract number, goods description, quantity, unit price and total amount. Invoices should be raised by the supplier in the name of consignee with their GSTIN only.
- (ii) Consignee Receipt Certificate as per attached format "**Annexure 'I'**" in original issued by the authorized representative of the consignee;
- (iii) Two copies of packing list identifying contents of each package;
- (iv) Inspection certificate issued by the nominated Inspection agency, if any;
- (v) Manufacturer 's/Supplier 's warranty certificate
- (vi) Certificate of origin (only in case of goods of foreign origin located in India).

b) On Acceptance:

Balance 20 % payment would be paid against "**Final Acceptance Certificate**" as per attached format "**Annexure 'J'**" of equipment's to be issued by the consignees subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise.

5. **Warranty: Rates should be quoted with comprehensive warranty of 3 years.** The operation, maintenance, and defect rectification of lab equipment shall be the responsibility of the supplier for a minimum period of three years from the date of operation of the lab. The bidders should attach duly signed and stamped certificate of **warranty** as per **Annexure-G** with the technical bid.
6. **Training:** In house training (where applicable) after the installation and commissioning of Equipment / Machinery/Instrument etc. shall be provided by the supplier.
7. **Delivery:** Delivery date will be mentioned in the supply order. The time and date of delivery or dispatch stipulated in a supply order shall be deemed to be the essence of the supply order and if the supplier fails to deliver or dispatch any consignment within the period prescribed for

such delivery, the delayed consignment will be accepted subject to penalty as laid down in the supply order, which will be recovered from the pending payments.

- i) No recovery of penalty will be made, if the delayed supplies are acceptable by extending the delivery period by Himachal Pharma Testing Lab Limited with our any Liquidated Damages.
- ii) Himachal Pharma Testing Lab Limited will allow extension on the request of the supplier by recording in writing that in exceptional circumstances the supply was beyond the control of the supplier and there was no loss to HPTLL.
- iii) Penalty on account of delay, Himachal Pharma Testing Lab Limited reserves the right to impose 0.5% (Half) per cent penalty per week on account of delay in supply, if delivery is received after expiry of the original delivery period. The total penalty will not exceed 10% of the value of the delayed goods.

8. Installation: Supplier has to install the Equipment / Machinery/ Instrument within two to three weeks from the receipt of the Equipment / Machinery/Instrument etc. in Himachal Pharma Testing Lab Limited, Baddi, Distt Solan H.P.

9. Site Preparation: The supplier shall inform Himachal Pharma Testing Lab Limited about the site preparation, if any, needed for the installation, immediately after receipt of the supply order. Supplier must provide complete details regarding space and all infrastructural requirements needed for the Equipment / Machinery/Instrument etc which Himachal Pharma Testing Lab Limited should arrange before the arrival of Equipment / Machinery/Instrument etc to ensure its early installation and smooth operation thereafter. The supplier may offer his advice and render assistance to Himachal Pharma Testing Lab Limited in the preparation of the site and other pre installation requirements.

10. The total scope of work includes the supply, installation, satisfactory commissioning and testing of the Equipment / Machinery/Instrument etc by the supplier, training at Himachal Pharma Testing Lab Limited, method development and validation for parameters as mentioned in specifications at **Annexure-A**.

11. Details about the service centre for the quoted Equipment / Machinery/Instrument etc. in India may be mentioned.

Read and Accepted.

(Signature & Stamp of Tenderer)

Acceptance

We _____ read and accept the instructions to the tenderer; terms & conditions and all other documents as mentioned in the tender and shall comply with them strictly.

Name of Bidder _____ Signature

Address _____ Seal of firm:

Date:

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

PERFORMA FOR PERFORMANCE STATEMENT

(For the Period of last 3 years)

Tender Notice No.						
Sr. No.	Order placed (Full) address of purchaser)	Order No. & date	Description And Qty. of Stores Ordered	Value of Order	Date of Completion of delivery	Has the Equipment/Item been satisfactory commissione d

Note: Copy of Completion certificate and Purchase Order shall be required.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

ANNEXURE: A (SPECIFICATION OF THE EQUIPMENT)

Sr. No.	Tender ID	Item No.	Description	Qty.
1	HPTLL_2025_03	HPTLL_2025_03_01	Gas Chromatography-Mass Spectrometry System	1
2		HPTLL_2025_03_02	Ultraviolet/ Visible Spectrophotometer	2

ITEM NO. HPTLL_2025_003_01

Gas Chromatography-Mass Spectrometry System

Technical specifications:

Basic Design and System Application:

· The system should have computer controlled fully automatic integrated Gas Chromatography-Mass Spectrometry System; Suitable for analysis of Nitrosamine impurities, Pesticides residues and other organic impurities in ppm, ppb and ppt levels with required sensitivity and stability of diverse range of drug substances, drug products, food products, herbal products and various types of water samples.

· The system should be a space saving, compact model that can fit into allocated lab space with all the sub- systems and accessories.

· Latest technology triple quadrupole GC-MS/MS with Combined/dedicated or equivalent Ionization source Electron Impact EI and **CI** sources to be provided enable to analyse Pharmaceuticals, Foods, and Water samples.

· Instrument will be utilized as per national and international regulatory guidelines.

· Corrosion-resistant exteriors should be provided.

· Model number of the equipment proposed to be supplied to be clearly mentioned.

· Gas Chromatograph should have self-diagnostic functions, built in keyboard and screen to control various parameters, apart from software controlling system through PC.

Compliance of system for measurement techniques:

Compliance of system for measurement techniques It must include a Gas Chromatograph with two injection ports, ECD detector, and Triple Quadrupole mass spectrometer with electron impact EI, **CI** & NCI, with liquid Auto-Injector and HS Sampler and PC workstation with licensed software. System should have a facility to run MS and ECD with single sample injection with variable split ratio for feed to MSD and ECD.

Gas Chromatography System:

Oven and column compartment:

· Temperature Range: **Ambient +2°C** to 450 °C or better

	· Temperature is directly selectable in 1 °C steps throughout the temperature range
	· Average ambient 0.05 °C change over a 5 °C change in ambient temperature, within the 10 to 35 °C sensitivity allowable ranges
	· Suitable for 0.53 mm diameter or equivalent size 2 columns.
Carrier-gas pneumatics:	
	· Carrier-gas pneumatics are suitable for 4 lines of gas and auto change over provision for carrier gas one to another.
	· Carrier-gas pneumatics are included with the injector and Head Space
	· Carrier PPC zones compensate for variations in ambient temperature and pressure for maximum stability
	· Split-vent pneumatics are included with the split/splitless and PSS (programmable split/splitless) injectors
	· PPC provides direct setting of split-flow rates and ratios
stability	· Split-vent PPC zones compensate for variations in ambient temperature for maximum stability
	· PPC provides direct setting in mL/min, psig or kPa or cm/sec
	· Automatic leak testing with PPC
	· Three-ramps pressure program
	· Pneumatic program rates: 0-100.0 psi/min, 0-100.0 mL/min, 0-200.0 cm/sec or ballistic
Auto-sampler:	
	· Two injector ports for capillary columns.
	· Injection speed Normal, fast, slow
	· Sample positions 100 or more, for 2-mL vials with wash vials 4 waste and 4 wash Waste and wash vial size 4 mL
	· Syringe size 0.5/ 1 µL, 5.0 µL or 50.0 µL injection volume 0.1 µL to 50.0 µL
	· Reproducibility < 0.5% RSD with 1 µL injection or better
	· Split/splitless capillary injector - 1 No. and Programmable Temperature Vaporizer - 1 No. , Split ratio easily adjustable for a wide range of analytical conditions
	· Injector temperature 50 °C to 450 °C in 1 °C increments or better
	· Minimum 4 software configurable modes: programmed/ Constant flow, programmed/ Constant pressure, programmed/ Constant linear velocity or constant flow. Vacuum compensation software selectable.
	· Automatic control of split vent by split flow or split ratio
	· Temperature-programmable inlet and fast heating and cooling rate.
Headspace (HS) Sampler:	
	· Minimum 80 vials of 20 ml,
	· Sample temperature settable from 35 °C to 300 °C in 1 °C increments or better

·	Needle temperature settable from 35 °C to 300 °C in 1 °C increments or better
·	Transfer line Temperature settable from 35 °C to 300 °C in 1 °C increments or better
·	Sampling method must be in such way that can produce the high degree of precision in replicate injection, RSD less than 5% or better with 1 mL injection or more.
·	Controls with inbuilt any controlling interface and with software same as GC
·	A standard size of touch screen with required keys in English for real-time graphical display of chromatogram and graphical display of temperature and pneumatic programs
Detector: Electron Capture Detector (ECD)	
·	Electron Capture Detector (ECD) with High sensitivity and selectivity with PPC pneumatics – software flow control of makeup gas
·	Source 15 mCi 63Ni or equivalent
·	Temperature Protect 470 °C by software
·	Makeup Gas Either Ar/CH4 or N2
·	Operating Temperature up to 400°C in 1 °C increments
·	Minimum Detectable Quantity < 5 fg lindane with argon/methane or nitrogen. (Preliminary specifications. Potential improvement under evaluation)
·	Linearity > 10 ⁴
height	· Data Acquisition Up to 500 Hz accommodate peaks as narrow as 5 m sec at half
MS-MS Detector:	
Ionization source:	
·	Combined/dedicated or equivalent electron impact EI and CI sources to be provided, with facility of interchanging easily by the user, and auto-detection of installed source by the instrument and software.
·	The ionization must be done both in a positive & negative mode.
·	The combined/dedicated ionization (EI, CI) source must operate along with reference spray to facilitate automated accurate mass measurements within single GC-MS experiment.
·	The instrument should be capable of internal reference mass correction for MS and MS/MS operation without losing sensitivity.
·	Automatic switching between EI and CI
·	The source should be easily removable from the system to facilitate user cleaning without venting the vacuum, with automatic standby of system while the source / probe is being removed
·	Electron Ionization (EI) source with maximum temperature of 350°C or better
·	All source parameters to be adjustable through software.
·	Electron energy up to 150 eV or more.
·	System should have dual filament with automatic filament switching.
Triple-Quadrupole & Ion Detector:	

·	A quadrupole mass analyser to provide effective ion transmission, superior resolution and abundance sensitivity
·	Mass Range: $\leq 10 - 1000$ or more (Da, amu)
·	Mass Analyser-Metal Quadrupole mass filter with pre rods
·	Mass Resolution-minimum 0.7 amu or better
·	Mass stability: Less than 0.1 Da over a 48 hour period
·	GC transfer line (interface) temperature Settable up to 350°C
·	Scan rate- Should have the scan speed of 10,000 amu/sec or better with fully variable
·	MRM speed (transitions/sec) 500 MRM/Sec or more
·	Simultaneous analysis of SRM or MRM with full SCAN/SIM and SCAN in one injection
·	Scan Modes: Precursor, Product, Neutral Loss, SIM, MRM
·	Sealed long-life electron multiplier
·	Lens for noise reduction and 10kv Conversion dynode
·	Must have wide dynamic range
·	Detection limits: Should be 0.5 fg or less (Proof of Statement must be provided)
·	Sensitivity: MRM EI: 40,000:1 for 100 fg/ μl of OFN or better (Proof of Statement must be provided)
·	Documentary evidence to be submitted along with quotation. For ten injections, %RSD should be
Vacuum System:	
·	System should have high capacity differential pumped vacuum system
·	Pump Air-cooled turbo molecular of 300 L/sec or more with separate pumping for ion source and quadrupole for homogenous vacuum backed up by rotary pump of at least 30ltr/min
·	Vacuum gauge provided as Standard for wide-range of vacuum monitoring.
·	Vacuum control should be fully automatic to make ON/OFF of each pump in startup & shut down, opening & closing of valve.
System Controller and Operating System:	
·	Software 21 CFR part 11 compliance, control for automatic data acquisition and processing.
·	Licensed software compatible to windows 11 OS (or downgraded to windows 10) to be provided with additional licenses with lifetime validity should be provided; as and when, there is a change in software upgradation it has to be provided without any additional charges.
·	Compatibility, connectivity, data acquisition and instrument controls provision with other software on networking server; control licence and drivers to be provided when required.
·	Mass spectrometer tuning and calibration automatic.
·	Data Validation (IQ/OQ/PQ for Software)
·	Self-diagnostics

· Remote diagnostics
· Software should control GC, Auto-sample, Headspace (HS) Sampler, MS and other accessories.
21 CFR Part 11 compliant through client server, for centralized safe and secure data management, project management, Applications and user Management etc.
Vendor to quote following Original, Licensed & latest version of Library/database with OEM part/catalogue number: NIST 2023 or latest and MRM database for Pesticides with 2000 optimized MRM transitions or more
Capability and critical control points:
· Instrument should be capable to analyse Nitrosamine impurities in drug substances and drug products, Pesticides residues and other organic impurities analysis in food products, herbal products and various types of water samples (but not limited to these analysis) at a concentration of ppb level with RSD of < 5% at standard conditions.
· Acquired data must compliance with the higher level of the data security as per 21 CFR part 11.
Standards:
· Specially pure Analytical NIST traceable standards with certificates for qualification of the whole system (GC & MS).
· Nitrosamine Standards: N-Nitrosodibutylamine, N-Nitrosodiethylamine, N-Nitrosodiisopropylamine, N-Nitrosodimethylamine, N-Nitrosoethylisopropylamine, N-Nitrosomethylaminobutyric Acid, N-Nitrosomethylphenylamine
· Pesticide standards: Organo Chloro Pesticides, Organo Phosphorous Pesticides, Herbicides Glyphosate and Glufosinate
Process specific requirement:
· Instrument shall be used for quantitative and qualitative analysis of pharmaceutical Raw Materials, Products, Food products and various types of water samples by different type of modes of operation to analyse the Nitrosamine impurities as per IP, USP, Ph. Eur. BP; and Pesticides residues & other organic impurities analysis in food products, herbal products and water samples as per EPA, BIS, IS and FSSAI.
· Instrument should be qualify the performance parameters as per pharmacopoeial requirement USP, BP, Ph. Eur., IP
· Auto detection and recognition of accessories and spares when attached with instrument.
· PC window based instrument controlling and data acquisition software.
· A complete library shall be available within the software for Nitrosamine impurities analysis as per IP, USP, Ph. Eur. BP; and Pesticides residues analysis in food products, herbal products and water samples as per EPA, BIS, IS and FSSAI.
· Language of software or firmware: English Language
Additional Accessories required for Installation of the System/Technique:

· All necessary pre-installation requisites/consumables including chemicals, acids, gases and standards for complete installation and demonstration of the instrument need to be supplied.
· Gas cylinders with regulators Helium, Nitrogen, Oxygen, Hydrogen as per system requirement to comply with all applications and regulations.
· Gas Panel as per requirement.
· Supplier must have proven capability and trained manpower to troubleshoot equipment both in terms of hardware and software.
Additional Consumables & Spares:
· The supplier of the instrument must confirm in writing that spares for the entire instrument (including additional units and accessories) will be available for at least a period 10 years after the model of equipment supplied has been phased out, for frequently required spares, there should be adequate inventory available with the Indian agency of the company.
· Suitable and essential tool kit is to be supplied with the instrument for required maintenance
a) Injection Port Septa – 300 Nos. b) Glass Insert with wool for Split & Splitless (10 Nos. Each) c) Vespel Ferrule 0.4 mm – 20 Nos., Vespel Ferrule 0.5 mm – 10 Nos. d) Graphite Ferrule 0.5 mm – 20 Nos., Graphite Ferrule 0.8 mm – 10 Nos. e) Auto-sampler Syringes 10 uL – 10 Nos. f) Auto-sampler Vials, caps & Septa - 1000 Nos. g) Filaments – 5 Nos., h) Glass Wool – 1 No., Column Cutter – 1 No. i) O-Ring – 20 Nos., Column Nut for GC and MS - 20 Nos. each j) Ion source cleaning kit – 1 No., GC-MS tool kit – 1 No. k) Triple Gas Filter Kit – 2 Nos.
Training & Support:
· The supplier will have to carry out successful installation at our laboratory premises (where ever the system has to be installed) and provide on-site comprehensive training for scientific personnel operating the system and support services till customer satisfaction with the system and a training at the suppliers lab premises is also required
· Supplier must have proven capability and trained manpower to troubleshoot equipment both in terms of hardware and software.
· List of accessories that are needed for the operation of the instrument, but not mentioned in the Technical Specifications.
· Documentary evidence must be furnished (client's certificate at least one from the Government sector) in support of satisfactory operation of the instrument; where the quoted model of equivalent sensitivity has already been installed, indicating LOD/LOQ of above mentioned elemental impurities relevant to Pharma & Food sector
· CMC with all consumable is included for 5 Years; Annual calibration of the equipment shall be a part of the CMC. It shall also be mandatory to perform calibration after every major repair/breakdown.
Safety requirements: Safety requirement with respect to man: No open wiring, no radiation leakage, Include warning systems.
Required Documents:
Installation qualification, operation qualification and performance qualification documents; etc.

Operational Manual of Hardware and software.
Test certificate of main equipments to be required
Deriver Qualification documents related to software.
Site area environmental conditions: Instrument installation and operational environment will be ambient as per Indian climate conditions; If any special it has to be mentioned in quotation.
Utilities:
Electric supply availability for Instrument: 220 to 240 v 50Hz AC single phase common power supply is available for all instruments through online UPS; If any special it has to be mentioned in quotation.
a) Capillary columns Rxi-5SiIMS/DB-5MS 30 m x 0.25 mm ID x 0.25 um or equivalent, Stabilwax/DB-Wax 30 m x 0.32 mm ID x 1.0 um or equivalent and Rxi-624SiIMS/DB-624MS 30 m x 0.32 mm ID x 1.8 um or equivalent – 1 No. each b) Gas cylinders with dual stage regulator for Argon & Helium – 2 Nos. each c) Wall mount Gas purification & control system d) Branded i7 PC, 8 GB RAM, 512 GB SSD, 27” Monitor, Licensed Windows 10 64-bit and laserjet Printer e) 10 KVA online UPS with isolation transformer with 60 min back up for complete system
Futuristic Scope:
· The Instrument hardware and software to be supplied with such accessories or arrangements so that this can be attached or synchronized with other allied techniques. (i.e. Liquid Chromatography)
A complete list or note to be provided by the vendor, about other instruments or techniques, by which this instrument can be attached or synchronized and quote the price separately for that.
system must be upgraded to 21 CFR Part 11 compliant through client server, for centralized safe and secure data management, project management, Applications and user Management etc.
Tablet Hardness Tester
TECHNICAL SPECIFICATIONS
· Measures 5 parameters - Weight, Thickness, Diameter (Length), Width and Hardness
· Full color intuitive touch screen display
· Constant speed and constant force measuring principle
· Up to 1000 samples per test
· Tilt sensing device enables display orientation change in verification and calibration mode
· Jaw with dual groove (big and small)
· 3 level user management security
· Indicator for calibration due date
· Light and audio alarm indicator for Instrument status
· Programmable calibration and verification modes

· Auto / Manual mode with programmable delay time
· Adjustable delay between the samples
· Tablet manager tray for easy carrying of tablets from weighing scale
· User selectable templates for printing reports and graphs
· Method transfer between equipments via USB
· Protective cover to protect operator from inhaling tablet dust from tablets that 'explode'
· Embedded printer for printing test results, validation and calibration reports
<u>Specifications:</u>
Display :
· Display : Intuitive touch screen
· Size : 7" full color
Hardness :
Units : N, Sc, Kp, user defined factor Accuracy : Better than $\pm 1N$
Max. force : 800N
Thickness / Diameter:
· Units : mm/inch
· Range : 0 - 40 mm
· Resolution : 0.01 mm
Weight:
· Weight : External balance connectivity
(Sartorius / Mettler / Shimadzu / Contech)
Constant speed and constant force :
· Constant speed setting : 0.5mm to 3mm in
step of 0.1mm Constant force setting : 10kg~100kg /sec force
in step of 1
kg/sec
Adjustable delay between the samples : 1 to 99 sec
Jaw :
Jaw opening : 40 mm
: up to 65mm (optional)
Jaw with groove : Dual (Big groove and small groove)
Hardness / Thickness Calibration :

Calibration : 2-point calibration Verification : 3-point linearity
Sample size : up to 1000
Interface : LAN, USB and serial
Printer connectivity : Embedded 40 column thermal printer, LAN and serial port
Standard accessories
Brush Standard accessories 01 NOS
Slide-in groove tongue* 01 NOS
Validation kit includes:
5 kg brass weight for calibration and validation of hardness
10 mm calibrated gauge block for validation of thickness
65 mm jaw
Flat tongue
Groove tongue
Magnetic tablet manager tray
STARLITE-12 Thermal paper roll (Set of 24)
Specifications :
· Image durability: 12 years *
· Sensitivity : high
· Protection Coating : topcoat
· Stability: Excellent stability against environmental influences
· Printability: Excellent printability of thermal coating side
· Roll length : 7 m
* specific handling and storage conditions have to be observed
IQ/OQ documents(SAC CODE-998346)

		Ultraviolet/ Visible Spectrophotometer
Sr. No.	Microprocessor based UV-Vis Spectrophotometer with the following specifications:	
1	Photometric System	Double beam optics
2	Photometric range	Absorbance: -4 to +4.0 Abs, Transmittance:0.0 to 400%
3	Photometric Accuracy	+/- 0.0015 Abs. at 0.5 Abs or better
4	Wavelength Range	190 to 1100 nm or better
5	Wavelength Accuracy	+/- 0.05 nm at D2 peak 656.1 nm
		± 0.3 nm or less in entire range
6	Wavelength Repeatability	+/- 0.025 nm or better
7	Photometric Repeatability	< +/- 0.00005 Abs at 0.5 Abs or lesser
		< +/- 0.00005 Abs at 1 Abs or lesser
		< +/- 0.00008 Abs at 2 Abs or lesser
8	Scanning speed	Up to 25,000 nm/min or better
9	Spectral Bandwidth	1 nm or better
10	Stray light	< 0.004 % at 220 nm (NaI) or lesser
		< 0.004 % at 340 nm (NaNO ₂) or lesser
		< 0.2% at 198 nm (KCl) or lesser
11	Baseline Stability	less than 0.0002 Abs/H
12	Baseline Flatness	less than 0.0003 Abs/H
13	Noise Level	Less than 0.00001 Abs
14	Monochromator	Czerny Turner blazed holographic grating
15	Detector	Silicone photodiode
16	USB Port connectivity	2-3 USB ports for data Transfer, PC, Printer
17	Light source	Tungsten and Deuterium lamp.
18	Quartz Cuvette	3 mL capacity, Pathlength of 10 mm (02 Pairs), 50 mm (1 Pair)
19	Spectrophotometer should have built in hardware validation for Wavelength accuracy, wavelength repeatability, resolution, stray light, photometric accuracy, photometric repeatability, baseline flatness, baseline stability, noise level.	
20	Windows based Operating software should have built in features like real time concentration display, Photometric mode Single / multi-wavelength, Enzyme Kinetics calculation, event recording such as Addition of reagents during measurement,	

	DNA/protein quantification etc. Spectrophotometer should have built-in color touch screen display so that it can be used without PC/Laptop also. Software must be 21 CFR Part 11 compliant.
21	Long pathlength cell holder for 10 to 100 mm pathlength for sample and reference.
22	Branded PC Core i7, 8 GB RAM, 512 GB SSD, 24" LED Monitor, Windows 10/11 Professional, optical mouse, keyboard, laserjet Printer, Compatible online UPS with isolation transformer with 60 minutes back up
23	Warranty: three years
24	Note: Vendor should clearly mention make and model, also enclose Catalogue/Data sheet of the Model quoted in support of the above mentioned Specifications otherwise their bids will be rejected.
25	Bidder should confirm compliance to following International safety guidelines and should enclose certificates alongwith bid from Principal company/Manufacturer:
	Electrical Safety (Low voltage directive (LVD) 2014/35/EU EN 61010-1:2010/A1:2019, EN 61010-2-061:2015), Electromagnetic Compatibility (EMC directive 2014/30/EU, EN 61326-1:2013), Restriction of Hazardous Substances (ROHS directive 2011/65/EU), Quality/Manufacturing Standards (Quality management system: ISO 9001:2015; Environmental management system: ISO 14001:2015.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure: "B"

FORMAT FOR PERFORMANCE BOND/GUARANTEE

(Undertaking from the supplier on a Non-Judicial Stamp Paper of requisite duly attested by Notary)

In consideration for "Himachal Pharma Testing Lab Limited" (hereinafter called HPTLL) having agreed to release the payment of net value as per terms and conditions of a concluded Order No. _____ dated _____ (**hereinafter called 'the order'**) for supply of _____ (**here in after called 'the Equipment /Item'**) to us Messrs _____ (**hereinafter called 'the supplier'**) on submission of a Performance Bond to the satisfaction of HPTLL for the due performance of the said order.

We, Messer's _____ hereby submit the FDR/DD No. _____ issued by _____ (Name of Bank) for _____ pledged in favour of Registrar, HPTLL as performance guarantee amount and hereby irrevocably, unconditionally and absolutely undertake against any loss or damage caused or suffered by HPTLL by reason of any failure of the supplier to perform or omission or negligence to perform any part of its obligations to the satisfaction of HPTLL in terms of the order.

We, the supplier, do hereby authorize HPTLL to forfeit this Performance Guarantee amount / undertake to pay the amount due and payable under this guarantee without any demur merely on a demand from the HPTLL stating that the amount claimed is due by way of loss or damage caused to or would be caused to or suffered by the HPTLL by reason of any breach by us of any of the terms and conditions contained in the said order or by reason of our failure or omission or negligence to perform the said order or any part thereof. We, the Supplier, undertake to pay to HPTLL any amount so demanded by HPTLL, notwithstanding:

- a) Any dispute or difference between HPTLL and supplier or any other person or between the supplier or any person or any suit or proceeding pending before any court or tribunal or arbitrator relating thereto; or
- b) The invalidity, irregularity or unenforceability of the order; or
- c) Any other circumstances which might otherwise constitute discharge of this guarantee, including any act of omission or commission on the part of HPTLL to enforce the obligations by the supplier or any other person for any reason whatsoever.

We, the Supplier, further agree that the performance Bond/ Guarantee herein contained shall be continued one and remain in full force and effect during the period that would be taken for the performance of the said order and that it shall continue to be enforceable till all the dues of the Himachal Pharma Testing Lab Limited under or by virtue of the said order have been fully paid and its claims satisfied or discharged or till the office of the CEO, Himachal Pharma Testing Lab Limited certifies that terms and conditions of the said order have been fully and promptly carried out by us and accordingly discharges this Performance Bond/ Guarantee.

We, the Supplier, further agree with HPTLL, that HPTLL shall have the fullest liberty without our consent and without affecting in any manner our obligations hereunder to vary any of the terms and conditions of the said order or to extend time of performance by the said supplier from time to time or to postpone for any time or from time to time and of the powers exercisable by the HPTLL against the said supplier and forbear or enforce any of the terms and conditions relating to the order and we shall not be relieved from our liability by reason of any such variation or extension being granted to us or for any forbearance, act or omission on the part of HPTLL or any indulgence

by HPTLL to us or by any such matter or thing whatsoever which under the law relating to sureties would, but for this provision, have effect of so relieving us.

This Performance Bond/Guarantee will not be discharged due to the change in the constitution of the supplier. We, the Supplier, undertake not to revoke this Performance Bond / Guarantee except with the prior consent of HPTLL in writing.

The disputes relating to this Bank Performance Bond / Guarantee shall be resolved as per the terms and conditions of the order.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure “C”

FORMAT FOR PERFORMANCE BOND (BANK GUARANTEE)

In consideration for the CEO HIMACHAL PHARMA TESTING LAB LIMITED, (*hereinafter called HPTLL*)

having agreed to release the payment of net value as per terms and conditions of a concluded Order No.

_____ dated _____ (*hereinafter called ‘the order’*) for supply of _____ (*hereinafter called ‘the Equipment /Item’*) to Messrs_ (*hereinafter called ‘the supplier’*) on submission of a Bank Guarantee to the satisfaction of HPTLL for the due performance of the said order.

We, _____ (*hereinafter called ‘the Bank’*) at the request of the supplier do, as a primary obliger and not merely as surety, hereby irrevocably, unconditionally and absolutely undertake against any loss or damage caused or suffered by HPTLL by reason of any failure of the supplier to perform or omission or negligence to perform any part of its obligations to the satisfaction of HPTLL in terms of the order.

We, the Bank do hereby undertake to pay the amount due and payable under this guarantee without any demur merely on a demand from HPTLL stating that the amount claimed is due by way of loss or damage caused to or would be caused to or suffered by HPTLL by reason of any breach by the said supplier of any of the terms and conditions contained in a said order or any part thereof. Any such demand made on the Bank shall be conclusive as regards the amount due and payable by the bank under this guarantee, which shall not be considered as satisfied by any intermediate payment or satisfaction of any part of or obligation hereunder. However, our liability under this guarantee shall be restricted to an amount not exceeding

We, the Bank, undertake to pay to HPTLL any amount so demanded by HPTLL notwithstanding

a). Any dispute and difference between HPTLL and supplier or any other person or between the supplier or any person or any suit or proceeding pending before any court or tribunal or arbitrator relating thereto or

a). The invalidity, irregularity or unenforceability of the order or

b). Any other circumstances which might otherwise constitute discharge of this guarantee, including any act of omission or commission on the part of HPTLL to enforce the obligations by the supplier or any other person for any reason whatsoever.

We, the Bank, further agree that the guarantee herein contained shall continue and remain in full force and effect during the period that would be taken for the performance of the said order and that it shall continue to be enforceable till all the dues of HPTLL under or by virtue of the said order have been fully paid and its claims satisfied or discharged or till the office of the Registrar, HPTLL confirms that the terms and conditions of the said order have been fully and promptly carried out by the said supplier and accordingly discharge this guarantee.

We, the Bank, hereby agree and undertake that any claim which the bank may have against the supplier shall be subject to and subordinate to the prior payment and performance in full

of all the obligations of the bank hereunder and the bank will not, without prior written consent of HPTLL, exercise any legal rights or remedies of any kind in respect of any such payment or performance so long as the obligations of the bank hereunder remain owing and outstanding, regardless of the insolvency, liquidation or bankruptcy of the supplier or otherwise. We, the Bank, will not counter claim or set off against its liabilities to HPTLL hereunder any sum outstanding to the credit of HPTLL with it.

We, the Bank, further agree with HPTLL, that HPTLL shall have the fullest liberty without our consent and without affecting in any manner our obligations hereunder to vary any of the terms and conditions of the said order or to extend time of performance by the said supplier from time to time or to postpone for any time or from time to time and of the powers exercisable by the HPTLL against the said supplier and forbear or enforce any of the terms and conditions relating to the order and we shall not be relieved from our liability by reason of any such variation or extension being granted to the said supplier or for any forbearance, act or omission on the part of HPTLL or any indulgence by HPTLL to the said supplier or by any such matter or thing whatsoever which under the law relating to sureties would, but for this provision, have effect of so relieving us.

This guarantee will not be discharged due to the change in constitution of the Bank or the supplier.

We, the Bank, lastly undertake not to revoke this Guarantee during its currency except with the prior consent of HPTLL in writing.

The disputes relating to this Bank Guarantee shall be resolved as per the terms and conditions of the order.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure "D"

Tender Form (Technical Bid)

Tender Subject:

Tender No.

1	Name of Bidder	
2	Address of the Bidder	
3	PAN Number of the Bidder (To be supported with the relevant documents)	
4	GST No. of the Bidder (To be supported with the relevant documents)	
5	Name of the Proprietor /Partner/ Director/ Authorized Person of the Bidding Agency. • in case of a company Board Authority. ** in case of Partnership firm Authority letter from other Partners.	
6	Contact No.	
7	Email	
8	Experience Details (To be supported with the relevant documents)	
9	Tender Document Cost Details (In the form of DD in the name of Himachal Pharma Testing Lab Limited)	DD No. Dated Amount in Rs. Five Thousand Only.
10	EMD Details (In the form of DD in the name of Himachal Pharma Testing Lab Limited)	DD No. Dated Amount in Rs.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure "E"

FORMAT FOR MANUFACTURER'S AUTHORIZATION FORM

To,

THE CEO

Himachal Pharma Testing Lab Limited

Baddi, Solan H.P.

Sub.: Tender for "_____".

Dear Sir,

We, _____, who are established and reputed manufacturers of _____, having factory/office at _____, hereby authorize M/s _____ [name & address of agents/distributors] to bid, negotiate and conclude the order with you for the above goods manufactured by us.

We shall _____ remain responsible for the tender/Agreement negotiated by M/s _____, jointly and severally.

We hereby extend our full guarantee and warranty as per the terms and conditions of tender for the goods offered for supply against this invitation for bid by the above supplier.

***Specify in detail manufacturer's responsibilities the services to be rendered by**

M/s _____ are as under:

i) _____

ii) _____

[Specify the services to be rendered by the agent/distributor] In case duties of the agent/distributor are changed or agent/ distributor is changed it shall be obligatory on us to automatically transfer all the duties and obligations to the new Indian Agent failing which we will ipso-facto become liable for all acts of commission or omission on the part of new Indian Agent/ distributor.

Yours faithfully,

[Name & Signature] For and on behalf of M/s. _____ [Name of manufacturer]

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure "F"

DECLARATION REGARDING BLACKLISTING / DEBARRING FOR TAKING PART IN TENDER

Self-Attested

I / We _____ (Tenderer) hereby declare that the firm / agency namely M/s. _____ has not been blacklisted or debarred in the past by Union / State Government or organization from taking part in Government tenders in India.

Or

I / We _____ (Tenderer) hereby declare that the Firm / agency namely M/s. _____ was blacklisted or debarred by Union / State Government or any Organization from taking part in Government tenders for a period of

_____ years w.e.f. _____ to _____. The period is over on

_____ and now the firm/company is entitled to take part in Government tenders.

In case the above information is found false I / we are fully aware that the tender/ contract will be rejected / cancelled by Himachal Pharma Testing Lab Limited, and EMD / SD shall be forfeited.

DEPONENT

Attested:

(Stamp of Company with authorized sign)

Name _____ Address _____

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpharmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure: "G"

CERTIFICATE OF WARRANTY

- i) .I/We certify that the warranty shall be for a period of _____ years for _____ and starting from the date of satisfactory installation, commissioning and handing over of the Equipment /Item and of the works conducted therewith covered under the supply order in working order. During the **warranty period, I/we shall provide free "after sale service" and the replacement of any part(s)** of the Equipment /Item or rectification of defects of work of the Equipment /Item will be free of cost. The replacement of the parts shall be arranged by us, at our own cost and responsibility. We undertake that the above warranty shall begin only from the date of satisfactory and faultless functioning of the Equipment /Item for 60 days at HPTLL, Baddi premises. The benefit of change in dates of the warranty period shall be in the interest of the us/your organization.
- ii). During the warranty period, we shall provide at least once in a quarter for **preventive maintenance visits**.
- iii). Uptime Guarantee: During the warranty period, we will be responsible to maintain the Equipment /Item in good working conditions for a period of 03 years.
- a). All complaints will be attended by us within 2 working days of receipt of the complaint in our office.
In case there is delay of more than 2 days in attending to a complaint from our side then you can count the number of days in excess of the permissible response time in the downtime. The above said response time of 2 days for attending to a complaint by us will not be counted in the downtime.
- b). **Penalty:** We shall pay a penalty equivalent to **0.5 % of the FOB/CIF** value of the Equipment /Item for every week or part thereof delay in rectifying the defect.
- Note: The right to accept the reason (s) for delay and consider reduction or wave off the penalty for the same shall be at the sole discretion of HPTLL, Baddi HP**
- iv). We certify that the Equipment /Item being/ quoted is the latest model and that spares for the Equipment /Item will be available for a period of at least ten years and we also guarantee that we will keep the organization informed of any update of the Equipment /Item over a period of ten year.
- v). We guarantee that in case we fail to carry out the maintenance within the stipulated period HPTLL, Baddi reserves the right to get the maintenance work carried out at our risk, cost and responsibility after informing us. All the expenses including excess payment for repairs/maintenance shall be adjusted against the Performance Bank Guarantee. In case the expenses exceed the amount of Performance Bank Guarantee, the same shall be recoverable from us with/without interest in accordance with the circumstances.
- vi). We shall try to repair the Equipment /Item at HPTLL, Baddi premises itself. However, the Equipment /Item will be taken to our site on our own expenses in case it is not possible to repair the same at HPTLL, Baddi. We shall take the entire responsibility for the safe custody and transportation of the Equipment /Item taken out for repairs till the Equipment /Item is rehabilitated to the HPTLL, Baddi after repair Any loss of Equipment /Item or its accessories under its charge on account of theft, fire or any other reasons shall be at our sole risk and responsibility which will be compensated to HPTLL, Baddi for such losses at the FOB/CIF value for the damaged/lost Equipment /Item part, including accessories.
- vii). We undertake to perform calibration after every major repair/breakdown/taking the Equipment /Item for repair out of HPTLL, Baddi premises.
- viii). In case of extended warrantee, we undertake to carry out annual calibration of the Equipment /Item.
- ix). We guarantee that we will supply spare parts if and when required on agreed basis for an agreed price. The agreed basis could be an agreed discount on the published catalogue price.
- x). We guarantee to the effect that before going out of production of spare parts, we will give adequate advance notice to you so that you may undertake to procure the balance of the life time requirements of spare parts.
- xi). We guarantee the entire unit against defects of manufacture, workmanship and poor quality of components.

Signature & Seal of the Manufacturer/Tenderer

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBNDA Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure-“H”

Tender Form (Financial Bid)

Tender Subject:

Tender No.:

Name of Bidder:						
Address of the Bidder:						
Sr. No.	Item No.	Description of Item (with Make and Model)	Unit	Approx Qty	Estimated Rate per unit in Rs	Estimated amount in Rs

**Signature of Bidder with
Date & Seal**

Note: The rate quoted above including all applicable Taxes, Freight, Installation, Commissioning and all other applicable taxes.

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure-“I”

CONSIGNEE RECEIPT CERTIFICATE

(To be given by consignee's authorized representative)

The following store(s) has/have been received in good condition:

- 1) Contract No. & date : _____
- 2) Supplier's Name : _____
- 3) Consignee's Name & Address with
telephone No. & Fax No. : _____
- 4) Name of the item supplied : _____
- 5) Quantity Supplied : _____
- 6) Date of Receipt by the Consignee : _____
- 7) Name and designation of
Authorized Representative of
Consignee : _____
- 8) Signature of Authorized
Representative of Consignee with
date : _____
- 9) Seal of the Consignee : _____

HIMACHAL PHARMA TESTING LAB LIMITED

EPIP, Phase-1, Near BBND Office, Jharmajri, Tehsil Baddi, Solan H.P. 174103

E-mail; himpfarmalab@gmail.com, Website: www.hptll.in

CIN: U71200HP2023SGC010333

Annexure-“J”

Proforma of Final Acceptance Certificate by the Consignee

No _____ Date _____

To

M/s _____

Subject: Certificate of commissioning of equipment/plant.

This is to certify that the equipment(s)/plant(s) as detailed below has/have been received in good conditions along with all the standard and special accessories and a set of spares in accordance with the contract/technical specifications. The same has been installed and commissioned.

(a) Contract No _____ dated _____

(b) Description of the equipment(s)/plants: _____

(c) Equipment(s)/ plant(s) nos.: _____

(d) Quantity: _____

(e) **Bill of Loading/Air Way Bill/Railway**

Receipt/ Goods Consignment Note no _____ dated _____

(f) **Name of the vessel/Transporter:** _____

(g) **Name of the Consignee:** _____

(h) **Date of commissioning and proving test:** _____

Details of accessories/spares not yet supplied and recoveries to be made on that account.

Sl. No.	Description of Item	Quantity	Amount to be recovered
---------	---------------------	----------	------------------------

The proving test has been done to our entire satisfaction and operators have been trained to operate the equipment(s)/plant(s).

The supplier has fulfilled its contractual obligations satisfactorily ## or

The supplier has failed to fulfil its contractual obligations with regard to the following:

He has not adhered to the time schedule specified in the contract in dispatching the documents/drawing pursuant to Technical Specification.

He has not supervised the commissioning of the equipment(s)/plant(s) in time, i.e. within

the period specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment(s)/plant(s).

The supplier as specified in the contract has not done training of personnel.

The amount of recovery on account of failure of the supplier to meet his contractual obligations is Rs----- (here indicate the amount).

Signature

Name

Designation with stamp

Explanatory notes for filling up the certificate:

He has adhered to the time schedule specified in the contract in dispatching the documents/drawings pursuant to "Technical Specification"

He has supervised the commissioning of the equipment(s)/plant(s) in time, i.e. within the time specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment(s)/plant(s).

Training of personnel has been done by the supplier as specified in the contract

In the event of documents/drawings having not been supplied or installation and commissioning of the equipment(s)/plant(s) having been delayed on account of the supplier, the extent of delay should always be mentioned in clear terms.